


Notes:

- Hydrophone Specifications:  
 Supply Voltage: 9V-15V  
 High Pass Cut Off: 1kHz  
 Sensitivity: -165dB re 1uPa  
 Configuration: Current Mode
- Potted junction to be cylindrical. Splice should be made as small as possible.
- Label Array: WHOI 214002
- Length Ordering Information: 214002-A-B-C-D-E

 WOODS HOLE OCEANOGRAPHIC INSTITUTION 1930	<b>Woods Hole Oceanographic Institution</b>	
	86 Water Street Dept 4 MS 18 Woods Hole, MA 02543	
	Title: <b>4 Element Hydrophone Array HTI-96-MIN (CM)</b>	
	Cage Code: <b>88846</b>	Drawing Number: <b>214002-SCH</b>
Drawn By: <b>Keenan Ball</b>	Date Drawn: <b>1/5/2009</b>	
Sheet 1 of 1		